Thermometer Calibration Log

Date	Thermometer location	How it was verified	Adjustment made	Employee

Using an insulated container add finely crushed ice and cold tap water, fully submerge thermometers indicator (at least 2 inches). The thermometer should read 32 degrees. If not adjust calibration nut, and check thermometer at least 30 seconds later. Continue to adjust as necessary.

Can also use a deep pan containing clean tap water bring to a full rolling boil. Submerge the thermometer indicator into the water. The thermometer should read 212 degrees. Make adjustments as necessary. Use proper precautions using this method.